

CERTIFICATE OF COMPLIANCE

MRI STRETCHERS

Common Name: MRI Stretchers

Trade Name and Model Number: Biodex MRI Stretcher, 240-110
 Biodex MRI Stretcher, Fowler Positioning, 240-100

The stretchers were tested based upon ASTM F2052-21, *Standard Test Method For Measurement Of Magnetically Induced Displacement Force On Medical Devices In The Magnetic Resonance Environment* with the following systems:

- 1.5 Tesla GE Signa LX MRI (HDx)
- 3.0 Tesla Siemens Magnetom Verio (GmbH)
- 7.0 Tesla Achieva MRI (Phillips)

Mirion Technologies (Capintec), Inc. certifies that the MRI stretchers referenced above were found to be MRI Conditional at 1.5 T, 3.0 T, and 7.0 T under the noted conditions.

Approved by



Mary Anne Yusko
Director, Regulatory Affairs

January 20, 2024

